

Übereinstimmung mit den Anforderungen zur CE-Erklärung

Hersteller: Hoover Dam Technology Asia Limited
Office B2, 16/F., Legend Tower
7 Shing Yip Street, Kwun Tong, Kowloon, Hong Kong

Produkt Name: Elektrischer Tester
Modell-Nummer: 3020, 3025

Hiermit bestätigen wir, dass die oben genannten Produkte mit den nachfolgenden Europäischen Richtlinien konform sind:

Richtlinie	Angewendete Norm
EMV-Richtlinie 2014/30/EU	EN 61326-1:2013 EN 55011:2016 + A1:2017
Niederspannungs-Richtlinie 2014/35/EU	EN 61010-1:2010
RoHS-Richtlinie 2011/65/EU + delegierte Richtlinie (EU) 2015/863 der Kommission	IEC 62321:2008 IEC 62321-3-1:2013 IEC 62321-4:2013 + A1:2017 IEC 62321-5:2013

Die oben genannten Produkte entsprechen der RoHS-Richtlinie 2011/65/EU des Europäischen Parlaments und des Rates vom 08.06.2011 zur Beschränkung der Verwendung gefährlicher Stoffe in Elektro- und Elektronikgeräten und der delegierten Richtlinie (EU) 2015/863 der Kommission vom 31.03.2015.

Dabei handelt es sich namentlich um folgende Substanzen:

Blei (Pb)
Cadmium (Cd)
Quecksilber (Hg)
Hexavalentes Chrom (Cr)
Polybromierte Biphenyle (PBB)
Polybromierte Diphenylether (PentaBDE, OctaBDE, DecaBDE)
Di(2-ethylhexyl)phthalat (DEHP)
Butylbenzylphthalat (BBP)
Dibutylphthalat (DBP)
Diisobutylphthalat (DIBP)

Hoover Dam Technology Asia Limited bestätigt hiermit, dass oben genannte Produkte RoHS-konform produziert werden.

Accordance with requirements for CE declaration

Manufacturer: Hoover Dam Technology Asia Limited
Office B2, 16/F., Legend Tower
7 Shing Yip Street, Kwun Tong, Kowloon, Hong Kong

Product Name: Electrical tester
Model Number: 3020, 3025

We herewith confirm that the products described above comply with the following European Directives:

Directive	Applied Standard
EMC Directive 2014/30/EU	EN 61326-1:2013 EN 55011:2016 + A1:2017
Low Voltage Directive 2014/35/EU	EN 61010-1:2010
RoHS Directive 2011/65/EU + Commission delegated Directive (EU) 2015/863	IEC 62321:2008 IEC 62321-3-1:2013 IEC 62321-4:2013 + A1:2017 IEC 62321-5:2013

The products described above comply with the RoHS Directive 2011/65/EU of the European Parliament and the Council from 08/06/2011 on restriction of the use of certain hazardous substances in electrical and electronic appliances and with the restriction of the Commission delegated Directive (EU) 2015/863 from 31/03/2015.


Following substances namely are involved:

Lead (Pb)
Cadmium (Cd)
Mercury (Hg)
Hexavalent Chromium (Cr)
Polybrominated Biphenyls (PBB)
Polybrominated Diphenyl Ethers (PentaBDE, OctaBDE, DecaBDE)
Bis(2-ethylhexyl) phthalate (DEHP)
Butyl benzyl phthalate (BBP)
Dibutyl phthalate (DBP)
Diisobutyl phthalate (DIBP)

Hoover Dam Technology Asia Limited herewith declares that the above-named products are manufactured in compliance with RoHS.

Hong Kong, 04. October 2019

Ort/Location, Datum/Date


Holger Emberger,
Head of Operations and Product
Management



Firmenstempel/Company Stamp

Certificate of Compliance

Certificate: 2180735

Master Contract: 247088

Project: 2180735

Date Issued: October 19, 2009

Issued to: **HDT Hoover Dam Technology Asia Ltd**
Unit 2008, 20/F, Yen Sheng Centre 64 Hoi Yuen Road,
Kwun Tong, Kowloon
Hong Kong

Attention: Mr. Jason Zhou

The products listed below are eligible to bear the CSA Mark shown



Issued by: Justin Liu



Authorized by: Anthony Lee P. Eng
Certification Specialist - Asia

PRODUCTS

CLASS -3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT
CLASS -3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE

Electrical Tester, double-insulated, Model 3020, rated 120/240/277/480 Vac, 0-65 Hz, and 160/330/380/600Vdc, 3.5 mA maximum.

Note: The above model is Class II Measurement Category III equipment for indoor use in Pollution Degree 2, 5 to 40 deg C ambient, and up to 2000 m environment

APPLICABLE REQUIREMENTS

- CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
- UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements
- CAN/CSA-C22.2 No. 61010-031-07 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 31: Safety requirements for hand-held probe assemblies for electrical measurement and test
- UL 61010-031 (1st Edition - 2007) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test

Certificate: 2180735

Project: 2180735



Master Contract: 247088

Date Issued: October 19, 2009

IEC 61010-031 (Edition 1.1 2008-06) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test (used as a guide for the flexing/pull test only)

Supplement to Certificate of Compliance

Certificate: 2180735

Master Contract: 247088

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
2180735	October 19, 2009	Original Certification.



TEST REPORT FOR FCC 47 CFR PART 15 SUBPART B Radio frequency devices	
Report Reference No.:	CGZ3181016-02177-E-D
Date of issue	19 October 2018
Testing Laboratory Name.....:	CTS (NINGBO) TESTING SERVICE TECHNOLOGY CO., LTD.
Address	A101, No.65, Zhuji Highway, Tianhe District, Guangzhou, Guangdong, China
Testing location/ procedure.....:	Full application of Harmonised standards <input checked="" type="checkbox"/> Partial application of Harmonised standards <input type="checkbox"/> Other standard testing method <input type="checkbox"/>
Applicant's name	Hoover Dam Technology Asia Limited
Address	Office B2, 16/F., Legend Tower, 7 Shing Yip Street, Kwun Tong, Kowloon, Hong Kong
Test specification:	
Standard.....:	47 CFR PART 15 SUBPART B ANSI C63.4:2014
Test Report Form No.:	CTSEMC-1.0
TRF Originator.....:	CTS (NINGBO) TESTING SERVICE TECHNOLOGY CO., LTD.
Master TRF	Dated 2009-01
CTS (Ningbo) Testing Service Technology Co., Ltd. All rights reserved. This publication may be reproduced in whole or in part for non-commercial purposes as long as the CTS (NINGBO) TESTING SERVICE TECHNOLOGY CO., LTD. is acknowledged as copyright owner and source of the material. CTS (NINGBO) TESTING SERVICE TECHNOLOGY CO., LTD. takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.	
Test item description.....	Electrical Tester
Trade Mark	HDT
Manufacturer	Hoover Dam Technology Asia Limited
Model/Type reference	3020
Ratings	AC 480V, DC 600V, 0-60Hz
Result	PASSED

Compiled by:

Kate zhang / Fileadministrators

Supervised by:

Duke yang / Technique principal

Approved by:

Vincent yao / Manager

Copyright of this report is owned by Centre of Testing Service and may not be reproduced other than in full except with the written approval of the issuing Company.

CTS (Ningbo) Testing Service Technology Co., Ltd.

A101, No.65, Zhuji Highway, Tianhe District, Guangzhou, Guangdong, China

Tel: +86-20-85543113 (32 lines)

Fax: +86-20-38780406

Complaint line: +86-20-85533471

E-mail: cts@cts-lab.com.cn

See Reverse For Terms And Conditions of Service

